

Certificate of Analysis									
Company:	Life Arises Farm		Black Afcandy						
		Lot: HL1-PL7				Report Date: 11/27/2023			
		Matrix: Flower				Date Analyzed: 11/22/2023			
Customer ID: 221107-2		Date Sampled: N/A				Analyst: 011			
Grower License #: SCLT0076-01		Date Received: 11/7/2023				Report ID: C231107BU			
Cannabinoid Summary									
Cannabinoid Profile	LOQ (mg/g)	Concentration (mg/g)	Weight (%)		18.04%		0.07%		
CBDVA	0.0005	<loq< td=""><td><lod< td=""><th>1</th><td rowspan="2">Total THC</td><td></td><td>Total CBD</td><td></td></lod<></td></loq<>	<lod< td=""><th>1</th><td rowspan="2">Total THC</td><td></td><td>Total CBD</td><td></td></lod<>	1	Total THC		Total CBD		
CBDV	0.0012	<loq< td=""><td><loq< td=""><th></th><td></td><td>Total CDD</td><td></td></loq<></td></loq<>	<loq< td=""><th></th><td></td><td>Total CDD</td><td></td></loq<>				Total CDD		
CBDA	0.0008	0.85	0.09			-			
CBGA	0.0008	14.46	1.45			-		-	
CBG	0.0019	0.59	0.06		22.12%		0.3%		
CBD	0.0019	<loq< th=""><th><loq< th=""><th>1</th><th>22.1270</th><th></th></loq<></th></loq<>	<loq< th=""><th>1</th><th>22.1270</th><th></th></loq<>	1	22.1270				
тнсу	0.0021	<loq< th=""><th><loq< th=""><th>1</th><th>Total</th><th></th><th colspan="2" rowspan="2">Δ9-ТНС</th></loq<></th></loq<>	<loq< th=""><th>1</th><th>Total</th><th></th><th colspan="2" rowspan="2">Δ9-ТНС</th></loq<>	1	Total		Δ9-ТНС		
CBN	0.0013	<loq< th=""><th><loq< th=""><th></th><th>Cannabinoids</th><th></th></loq<></th></loq<>	<loq< th=""><th></th><th>Cannabinoids</th><th></th></loq<>		Cannabinoids				
Δ9-ТНС	0.0020	2.98	0.30					1	
Δ8-THC	0.0019	<loq< th=""><th><loq< th=""><th></th><th></th><th>_</th><th></th><th>_</th></loq<></th></loq<>	<loq< th=""><th></th><th></th><th>_</th><th></th><th>_</th></loq<>			_		_	
THC-A	0.0034	202.34	20.23		40.520/		1.0		
СВС	0.0024	<loq< th=""><th><loq< th=""><th>]</th><th>10.53%</th><th></th><th colspan="2">1:0</th></loq<></th></loq<>	<loq< th=""><th>]</th><th>10.53%</th><th></th><th colspan="2">1:0</th></loq<>]	10.53%		1:0		
Total THC		180.44	18.04	1	Percent		THC : CBD		
Total CBD		0.75	0.07		Moisture		Ratio		

22.12

Cannabinoids Methodology: High Performance Liquid Chromatography (HPLC) using PerkinElmer FLEXAR[™] with Photo Diode Array Detector (PDA)

221.23

Total CBD and total THC are calculated values, to account for assumed decarboxylation from the acid form (THCA or CBDA) to the neutral form, causing weight loss of the acid group. These values are calculated as follows: Total CBD = (CBDA x 0.877) + CBD Total THC = (THCA x 0.877) + Δ 9-THC Ratio of Total CBD: Total THC Reagent Blanks: < LOQs for all analytes

LOQ = The lowest quantity that this method can reliably detect. Any cannabinoid that was not detected is assumed to be less than the stated LOQ (<LOQ).

All results reflect dry weight of material, based on % moisture of the sample.

Measurement of Uncertainty (MU): the parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the particular quantity subject to measurement. Δ 9-THC MU = ±0.005% Total THC MU = ±0.007%

All other cannabinoid MU values are available upon request.

Total Cannabinoids

All moisture analysis is determined by loss-on-drying measurement using OHAUS Model MB90 Moisture Content Readers.

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the Certified by: samples as received.

ule E.M

SULT0076-01

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

C231107BU

BLACK AFCANDY

(802) 540-0148 laboratory@biadiagnostics.com Certificate Registration Number: CL_50_2021_002



Customer ID: 221107-2

Grower License #: SCLT0076-01

Company: Life Arises Farm

Certificate of Analysis

Sample ID: Black Afcandy Lot: HL1-PL7 Matrix: Flower Date Sampled: N/A Date Received: 11/7/2023

Report Date: 11/27/2023 Date Analyzed: 11/21/2023 Analyst: 048 Report ID: C231107BU

Water Activity Summary

Test	Method	Result	
Water Activity	ASTM D8196: Determination of Water Activity in Cannabis Flower	0.4255	



Test Methodology: Aqualab TDL 2 water activity meter with tunable diode laser

This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Luke E.M.

Luke Emerson Mason (Laboratory Director, Bia Diagnostics)

(802) 540-0148 laboratory@biadiagnostics.com

Certified by: